



CSA INTERNATIONAL

# Certificate of Compliance

**Certificate:** 1714272

**Master Contract:** 226885

**Project:** 2033264

**Date Issued:** 2008/05/21

**Issued to:** Daniel Chromatograph Division  
Daniel Measurement and Control Inc.  
11100 Brittmoore Park Drive  
Houston, TX 77041  
USA  
Attention: Harold Sanders

*The products listed below are eligible to bear the CSA Mark shown*



**Issued by:** Matthew Brock

**Authorized by:** Patricia Pasemko, Operations  
Manager

## **PRODUCTS**

**CLASS 2252 03 - PROCESS CONTROL EQUIPMENT**

Gas Chromatograph, Model: 2350A

Rated Input: 115/230 Vac, 50/60 Hz, 0.33A, All signal circuits SELV, Ambient -18 to 55 deg. C, Installation Category II, Pollution Degree 2.

## **APPLICABLE REQUIREMENTS**

- CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements.



CSA INTERNATIONAL

**Certificate:** 1714272

**Master Contract:** 226885

**Project:** 2033264

**Date Issued:** 2008/05/21

---

### **MARKINGS**

The following markings appear on the product and may be described in further detail in the Certification Report Supplement, Section C:

- Submitter's identification (company name and/or file number and/or registered tradename);
- Factory identifier;
- Model designation;
- Electrical rating;
- Date of manufacture (or traceable serial number);
- CSA Mark;
- CSA Enclosure Type;
- Terminal Identification;
- Maximum ambient.