



Supplement to Certificate of Compliance

Certificate: 1664959

Master Contract: 226885

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
1983994	2008/05/08	Update to report 1664959 to include new drawings, alternate components and report corrections.
2020306	2008/03/27	Revision to report 1664959, to Change FR-4 laminate (Polycalad FR-370HR to Polyimide laminate (Arlon35) on FID Connector board assembly.
1941983	2007/12/06	Revision to report #1664959, to add option liquid sample inject valve & flame ionization detector
1856459	2007/06/12	Update to report 1664959 to include retesting and addition of Division Explosion proof ratings (Class I, Div 1 / Div 2, Groups B, C, D) and the addition of an alternate LOI interface assembly on the Lower Electronics Enclosure.
1785439	2006/07/05	Update to report 1664959 to include optional power supply.
1789829	2006/05/17	Update to report 1664959 to include alternate enclosure.
1664959	2005/04/22	Gas Chromatograph, Model 700, Evaluation for Hazardous Locations, AEx/Ex d IIB+H2, T4, Enclosure Type 4

History

Testing was conducted through several previous CSA projects 1346684, 1469334, & 1575274.

All test data is brought forward to this project 1664959