

British Approvals Service for Electrical
Equipment in Flammable Atmospheres



Certificate of Conformity - Variation

SUPPLEMENTARY CERTIFICATE BAS No. Ex 96D2494/4

This is to certify that Apparatus Certificate number:

Ex 96D2494

held by:

ROSEMOUNT ANALYTICAL INC

of:

2400 Barranca Parkway, Irvine, California 92714-5018, USA

for the:

pH/ORP TRANSMITTER MODEL 3081


is hereby extended to apply to the apparatus designed and constructed in accordance with the specification set out in the Schedule of the said Certificate but having any variations specified in the Schedule attached to this certificate and the documents therein referred to.

This Supplementary Certificate shall be held with the original Certificate.

File No: EECS 0911/02/015

Sheet 1 of 2

This certificate is granted subject to the general conditions of the Electrical Equipment Certification Service. It does not necessarily indicate that the apparatus may be used in particular industries or circumstances. Representation of equipment as "Certified" is valid only when the number of the prime certificate to which this certificate is a supplement is given on the relevant EECS Manufacturing Licence or Verification Certificate.



I M CLEARE
DIRECTOR
27 July 1998



Registration Number
020
The use of the Accreditation
Mark indicates accreditation in
respect of those activities
covered by the accreditation
certificate number 020.


HSE
Health & Safety
Executive

Electrical Equipment Certification Service
Health and Safety Executive
Harpur Hill, Buxton, Derbyshire, SK17 9JN, United Kingdom
Tel: 01298 28000 Fax: 01298 28244



Supplementary Certificate BAS No. Ex 96D2494/4

VARIATION FOUR

To permit the following:-

1. Changes to the CPU printed circuit board which include:
 - a) Minor correction to pin designation on connector JP4
 - b) Minor changes to production details
 - c) Alternative pin configuration on various jumper connectors.
2. The use of an alternative software program in the previously certified models thus creating a new Model 3081FG.

The above changes do not affect intrinsic safety.

DRAWINGS

<u>Number</u>	<u>Issue</u>	<u>Date</u>	<u>Description</u>
23641-02/04	C	6.5.98	PCB assembly, CPU
9241337-00	A	Jul 13 98	Label - Model 3081 FG