



RECOMMENDED D > 150 [6].  
 MIN. D = 50 [2] REQUIRES SPECIAL CONFIG. SETTINGS.  
 MAX REC. H < (D + 100) [D + 4].

WIRE CLIP/CHUCK (OPTIONAL)

MIN 5X WIRE  
 MIN 3X WIRE  
 MIN 5X WIRE

DEPENDING ON TURBULENCE IN TANK AND MOUNTING ANGLE THE PROBE MAY NEED TO BE ANCHORED OR SUPPORTED.

FOR MORE INFO, SEE REFERENCE MANUAL.

- GUIDELINES AND TANK CONNECTIONS VALID FOR BOTH  $\phi 4$  AND  $\phi 6$  WIRE.
- P/T-RATING FOR FLANGE: SEE REF. MANUAL.

FOR MORE INFO, SEE REFERENCE MANUAL.

POS	ITEM DESCRIPTION	QTY	REMARK	SUPPLIER
04	GASKET	1	ACC. TO BSP-G THREAD STD.	CUSTOMER
03	GASKET	1	SOFT GASKET T=MIN 1.6 (ANSI B16.5/1A, EN1514-1/1BC) SEE NOTE 1	CUSTOMER
02	SCREW/NUT/WASHER	-	ACC. TO FLANGE STD. SEE NOTE 1.	CUSTOMER
01	FLANGE	1	DIM. ACC. TO ANSI B 16.5/ EN 1092-1, OTHER ON REQUEST. P/T-RATING SEE NOTE 1.	RMT
POS	ISSUED BY	WEEK	PRODUCT CODE	FILE
	GU-TK-RS	0311	3300	IDEAS
	GU-HF	0319	9150077-974	SCALE 1:4
	ROSEMOUNT			1 ST ANGLE
				DOC TYPE 2
				DWG NO. 9150077-974
				ISSUE 2
				SHEET 01/01